




America

CERTIFICATE

No. U8 071797 0006 Rev. 02

Holder of Certificate: **Vista Medical Ltd.**
 Unit #3-55 Henlow Bay
 Winnipeg MB R3Y 1G4
 CANADA

Certification Mark:



Product: **General Medical Devices
 Pressure Mat**

This product was voluntarily tested to the relevant safety requirements referenced on this certificate. It can be marked with the certification mark above. The mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL and a Standards Council of Canada accredited Certification body.

Test report no.: 7169005318-200

Date, 2020-12-17

(Vivek Chandna)



America

CERTIFICATE

No. U8 071797 0006 Rev. 02

Model(s): BT2-xyyy-zzz

Where “xx” and “yy” represent the dimensions of the array, “xx” and “yy” can be any numerical numbers ranging from 04 to 64. “zzz” represents style, “zzz” can be any numerical numbers ranging from 000 to 999

Tested according to:

CAN/CSA-C22.2 No. 60601-1:2014 excluding Biocompatibility (Cl. 11.7) and EMC (Cl. 17)
 CAN/CSA-C22.2 No. 60601-1-6:2011/A1:2015-10
 ANSI/AAMI ES60601-1:2005/A1:2012 excluding clauses 11.7 (Biocompatibility) and 17 (EMC)
 Supplemented by
 IEC 62304:2006/AMD1:2015
 IEC 62304:2006

Production Facility(ies): 071797

Parameters:

Rated Input Voltage: 5Vdc
 Rated Input Current: 0.5A
 Protection Class: II
 Degree of Protection: Type B applied part